

Tunable Diode Laser Summary Report

TERRA AIR SERVICES
 2950 North Loop West
 Suite 1050
 Houston, Texas 77092
 Phone (713) 956-6250
 FAX (713) 956-2185

Customer: Tosco SFR

Monitoring Period

Address: 1380 San Pablo Avenue
 Rodeo, CA 94572-1299

From: 10/01/1997 00:00:01 PST

To: 10/31/1997 23:59:59 PST

North Fenceline

Compound	Average Concentration (ppm)			Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#			
Ammonia	0.000	0.000	0.000	100	0	3	0	744	99.79	0.21
Hydrogen sulfide	0.036	3.438	0.000	15	0	0.03	161	744	98.81	1.19

South Fenceline

Compound	Average Concentration (ppm)			Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#			
Ammonia	0.000	0.000	0.000	100	0	3	0	744	92.19	7.81
Hydrogen sulfide	0.000	0.000	0.000	15	0	0.03	0	744	99.29	0.71