

TERRA AIR SERVICES
Ultra Violet
Summary Report

Customer: Tosco SFR
 Address: 1380 San Pablo Avenue
 Rodeo, CA 94572-1299

Monitoring Period
 From: 05/01/1998 00:00:01 PST
 To: 05/31/1998 23:59:59 PST

PETRIS
 TECHNOLOGY
 2950 N. Loop W., Ste. 1050
 Houston, Texas 77092
 Phone (713) 956-2165
 Fax (713) 956-2185

North Fenceline

Compound	Concentration* (ppm)			Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#			
Benzene	0.000	0.017	0.000	5	0	1	0	744	4.42	95.58
Carbon disulfide	0.004	0.099	0.000	10	0	1	0	744	78.70	21.30
m-Xylene	0.001	0.022	0.000	200	0	1	0	744	4.42	95.58
o-Xylene	0.001	0.001	0.000	200	0	1	0	744	78.70	21.30
p-Xylene	0.000	0.003	0.000	200	0	1	0	744	4.42	95.58
Sulfur dioxide	0.006	0.028	0.000	3	0	0.3	0	744	78.70	21.30
Toluene	0.000	0.010	0.000	100	0	50	0	744	4.42	95.58

South Fenceline

Compound	Concentration* (ppm)			Alarms				Total Hours	Online %	Offline %
	Avg	Max	Min	High	#	Low	#			
Benzene	0.004	0.018	0.000	5	0	1	0	744	22.50	77.50
Carbon disulfide	0.000	0.003	0.000	10	0	1	0	744	54.41	45.59
m-Xylene	0.000	0.000	0.000	200	0	1	0	744	22.50	77.50
o-Xylene	0.000	0.021	0.000	200	0	1	0	744	54.41	45.59
p-Xylene	0.000	0.013	0.000	200	0	1	0	744	22.50	77.50
Sulfur dioxide	0.003	0.044	0.000	3	0	0.3	0	744	54.41	45.59
Toluene	0.006	0.054	0.000	100	0	50	0	744	22.50	77.50

* Concentration Results based on 1 minute averages